

INTOX EC/IR II
Quality Assurance Worksheet

Handwritten initials: HNS and MSK

Instrument Serial Number 010468 Worksheet Start Date 10/17/2013

Location Fairfax County Adult Detention Center

Address 10520 Judicial Dr., Fairfax, VA 22030

DFS Technician Heather Stanton License No. 18910

☐ Laboratory ☒ On-Site

Site Specifications

Yes No

1 Secured Environment:

Location must not be accessible to the general public.

☒ ☐

2 Environment free from contaminants

☒ ☐

3 Dedicated telephone line:

The phone line should be used solely by the instrument.

☒ ☐

4 Adequate Ventilation:

Enough space must be provided for the instrument and the printer to allow for adequate ventilation.

☒ ☐

5 Stable temperature controls:

The instrument should not be exposed to direct sources of temperature variation.

☒ ☐

6 115VAC power source:

Whenever possible instrument power circuit should not be shared with refrigerators, copy machines or other high amperage devices.

☒ ☐

7 Room Condition:

The site should be free of excessive moisture, heat and dust.


☒ ☐

E-Seek
Keyboard
Printer
UPS

Old serial no.

New serial no.

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Maintenance (if any)

Comments (if any)

Instrument Serial Number 010468

Certification Date

Instrument Barometer (mm HG) 746

Reference Barometer (mm HG) 748

Reference Barometer(RB)Serial # 009113

RB Calibration Due 5/8/2014

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

all measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

☐ Calibrated☐ Certified☐ Replaced dry gas standard (+O-ring)

Dry gas standard Lot No. (with tank no.)

AG210901-09

☒ Measurement Assurance Check☒ Run a Breath Test (Instrument Test)

(successful completion denotes satisfactory condition of the item)

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.098
0.100	0.098	0.095	0.101	Sample 2	0.098
Precision		sample min	sample max	Sample 3	0.098
0		0.098	0.098		

DFS Technician

Heath A. Zisch

Date

10/18/13

Issuing Analyst

Melissa Kennedy

Date

10/19/13

Intox EC/IR II QA Worksheet

Issued by Calibration and Training Program Manager

Issue Date: 4-March-2013

DFS Document 250-F113

Revision 6

Page 2 of 2

Intox EC/IR-II: Accuracy Check

ANS

Department of Forensic Science

MSK

Serial Number: 010468 Test Number: 834

Test Date: 10/17/2013 Test Time: 09:13 EDT

Dry Gas Target: 0.295

Lot Number: AG322702-01 Exp Date: 08/15/2015

Tank Pressure: 497 psi Barometric Pressure: 746 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	09:14
CHK	0.299	09:15
BLK	0.000	09:17
CHK	0.297	09:17
BLK	0.000	09:19
CHK	0.297	09:20

Test Status: Success

Calibration CRC: 21A2382B

AKB
NSK

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010468 Test Number: 835

Test Date: 10/17/2013 Test Time: 09:22 EDT

Dry Gas Target: 0.098

Lot Number: AG210901-09 Exp Date: 04/18/2014

Tank Pressure: 482 psi Barometric Pressure: 746 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	09:22
CHK	0.098	09:23
BLK	0.000	09:25
CHK	0.098	09:25
BLK	0.000	09:27
CHK	0.098	09:28

Test Status: *Success*

Calibration CRC: 21A2382B

ANS
MSK




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 10/17/2013
TEST EQUIPMENT NUMBER 010468		

RESULTS: TIME SAMPLE TAKEN 10:03 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 18 Oct 2013 at 08:19

Handwritten initials: HES and MSK

Test Results

Instrument Serial Number 010468

Test # 000836 **Subject Test**

Test Location 1 Department of
Test Date 17 Oct 2013

Test Location 2 Forensic Science
Test Time 09:55
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M
License Number 18910

Card Serial Number 118910

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 17 Oct 2013 **End Time** 10:04

Result Time 10:03

Result Date 17 Oct 2013 **Result** 0.00

Data Type DIAG

Sample Value Pass

Sample Time 09:55

Data Type BLK

Sample Value 0.000

Sample Time 09:56

Data Type CHK

Sample Value 0.098

Sample Time 09:57

Data Type BLK

Sample Value 0.000

Sample Time 09:58

Data Type SUBJ

Sample Value 0.000

Sample Time 09:59

Data Type BLK

Sample Value 0.000

Sample Time 10:00

Data Type SUBJ

Sample Value 0.000

Sample Time 10:03

Data Type BLK

Sample Value 0.000

Sample Time 10:04

Standard Type Dry Gas Std

Standard Value 0.098

Standard Lot Number AG210901-09

Standard Expiration Date 04/18/2014

Tank Pressure 473

Barometric Pressure 746 mmHg

Blow Sample Number 1 **Blow Duration** 3.52 sec

Blow Volume 1697 cc **End-of-Blow Time** 09:59

Blow Sample Number 2 **Blow Duration** 3.41 sec

Blow Volume 1718 cc **End-of-Blow Time** 10:03

Tamper Evident Stamp c7baab3b

Test Status Code 0

Test Status Success